

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081919\
 Data File : VV012299.D
 Acq On : 20 Aug 2019 01:27
 Operator : SY/MD
 Sample : K4373-18
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AE4

Quant Time: Aug 20 04:42:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR081919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 20 02:35:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	243567	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	228110	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	117119	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	63523	5.10	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.00%
7) Chloroethane-d5	1.58	69	49317	5.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.40%
11) 1,1-Dichloroethene-d2	2.13	63	51243	2.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	43.40%#
20) 2-Butanone-d5	3.94	46	202231	60.27	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.54%
24) Chloroform-d	4.40	84	140550	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.60%
26) 1,2-Dichloroethane-d4	5.08	65	74204	4.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.60%
32) Benzene-d6	5.10	84	287428	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
36) 1,2-Dichloropropane-d6	6.12	67	88093	5.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.20%
41) Toluene-d8	7.36	98	254661	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	7.66	79	31073	4.24	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.80%
46) 2-Hexanone-d5	8.13	63	165955	71.77	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	143.54%#
57) 1,1,2,2-Tetrachloroethane-	10.26	84	68977	5.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	107993	4.58	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.60%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	2.19	43	119073	67.840	ug/L #	61
14) Carbon disulfide	2.32	76	23759	0.766	ug/L	98
21) 2-Butanone	4.03	43	34021	10.216	ug/L	99
33) Benzene	5.14	78	130375	2.066	ug/L	100
40) 4-Methyl-2-pentanone	7.27	43	14917	1.924	ug/L #	85
42) Toluene	7.43	91	32274	0.462	ug/L	99
52) Ethylbenzene	9.05	91	159550	2.042	ug/L	98
53) m,p-xylene	9.18	106	60707	1.990	ug/L	96
54) o-xylene	9.59	106	28623	0.976	ug/L	94
56) Isopropylbenzene	9.97	105	34922	0.431	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081919\
Data File : VV012299.D
Acq On : 20 Aug 2019 01:27
Operator : SY/MD
Sample : K4373-18
Misc : 25mL/MSVOA V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AE4

Quant Time: Aug 20 04:42:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR081919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 20 02:35:23 2019
Response via : Initial Calibration

