

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081920\
 Data File : VV018057.D
 Acq On : 19 Aug 2020 10:27
 Operator : SY/MD
 Sample : VV0819MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK85

Quant Time: Aug 20 04:36:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 20 04:32:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	281409	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	278719	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	132558	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	77333	36.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.04%
7) Chloroethane-d5	1.58	69	69660	39.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.32%
11) 1,1-Dichloroethene-d2	2.12	63	120868	29.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.86%#
21) 2-Butanone-d5	3.91	46	119350	84.87	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.87%
24) Chloroform-d	4.37	84	158524	40.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.12%
26) 1,2-Dichloroethane-d4	5.05	65	107518	42.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.70%
32) Benzene-d6	5.07	84	309367	40.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.56%
36) 1,2-Dichloropropane-d6	6.09	67	102308	40.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	81.48%
41) Toluene-d8	7.33	98	267010	38.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.30%#
43) trans-1,3-Dichloropropene-	7.64	79	45710	40.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.26%
47) 2-Hexanone-d5	8.11	63	57104	70.86	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	70.86%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	128071	39.16	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.32%
64) 1,2-Dichlorobenzene-d4	11.65	152	118889	45.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.14%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081920\
Data File : VV018057.D
Acq On : 19 Aug 2020 10:27
Operator : SY/MD
Sample : VV0819MBL01
Misc : 5.0µL/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK85

Quant Time: Aug 20 04:36:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 20 04:32:09 2020
Response via : Initial Calibration

