

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081920\
 Data File : VV018058.D
 Acq On : 19 Aug 2020 10:57
 Operator : SY/MD
 Sample : VV0819WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK86

Quant Time: Aug 20 04:41:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 20 04:32:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	272534	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	271517	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	131297	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	76921	36.99	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.98%
7) Chloroethane-d5	1.58	69	72045	42.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.70%
11) 1,1-Dichloroethene-d2	2.12	63	121581	30.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.14%
21) 2-Butanone-d5	3.91	46	123706	90.83	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.83%
24) Chloroform-d	4.37	84	162877	43.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.06%
26) 1,2-Dichloroethane-d4	5.05	65	112749	45.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.72%
32) Benzene-d6	5.07	84	319279	42.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.34%
36) 1,2-Dichloropropane-d6	6.09	67	108116	44.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.38%
41) Toluene-d8	7.33	98	275426	40.93	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.86%
43) trans-1,3-Dichloropropene-	7.64	79	48369	43.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.18%
47) 2-Hexanone-d5	8.11	63	59043	75.21	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.21%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	133684	41.97	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.94%
64) 1,2-Dichlorobenzene-d4	11.65	152	125505	48.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.14%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081920\
Data File : VV018058.D
Acq On : 19 Aug 2020 10:57
Operator : SY/MD
Sample : VV0819WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK86

Quant Time: Aug 20 04:41:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 20 04:32:09 2020
Response via : Initial Calibration

