

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081921\
 Data File : VV021887.D
 Acq On : 19 Aug 2021 13:35
 Operator : SY/MD
 Sample : M3422-09DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKB0DL

Manual Integrations
 APPROVED

MMDadoda
 8/20/2021 12:53:14 PM

Quant Time: Aug 20 02:17:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 20 02:14:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	461689	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	428984	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	231676	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	135191	37.456	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	74.920%		
7) Chloroethane-d5	1.561	69	107908	39.618	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	79.240%		
11) 1,1-Dichloroethene-d2	2.105	63	172810	30.079	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	60.160%		
21) 2-Butanone-d5	3.876	46	197271	87.880	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	87.880%		
24) Chloroform-d	4.352	84	254129	40.639	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	81.280%		
26) 1,2-Dichloroethane-d4	5.037	65	154011	42.817	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	85.640%		
32) Benzene-d6	5.053	84	539364	43.723	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	87.440%		
36) 1,2-Dichloropropane-d6	6.072	67	167516	43.606	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	87.220%		
41) Toluene-d8	7.320	98	499874	43.841	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	87.680%		
43) trans-1,3-Dichloroprop...	7.625	79	74036	39.160	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	78.320%		
47) 2-Hexanone-d5	8.088	63	127893	101.577	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	101.580%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	234069	49.168	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	98.340%		
66) 1,2-Dichlorobenzene-d4	11.625	152	212825	44.388	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	88.780%		
Target Compounds						Qvalue
18) Methyl tert-butyl Ether	2.777	73	10113m	1.027	ug/L	
33) Benzene	5.108	78	30683	2.324	ug/L	100
42) Toluene	7.391	91	159082	11.283	ug/L	97
52) Ethylbenzene	9.014	91	645294	42.900	ug/L	99
53) m,p-Xylene	9.140	106	126793	21.750	ug/L	100
54) o-Xylene	9.545	106	74844	12.952	ug/L	100
55) Styrene	9.564	104	89715	9.271	ug/L	100
62) 1,3,5-Trimethylbenzene	10.542	105	40604	3.115	ug/L	100
63) 1,2,4-Trimethylbenzene	10.918	105	131979	9.959	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081921\
Data File : VV021887.D
Acq On : 19 Aug 2021 13:35
Operator : SY/MD
Sample : M3422-09DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1


Instrument :
MSVOA_V
ClientSampleId :
DBKB0DL

Manual Integrations
APPROVED

MMDadoda
8/20/2021 12:53:14 PM

Quant Time: Aug 20 02:17:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 20 02:14:47 2021
Response via : Initial Calibration





#18





Methyl tert-butyl Ether
 Concen: 1.027 ug/L m
 RT: 2.777 min Scan# 540
 Delta R.T. 0.013 min
 Lab File: VV021887.D
 Acq: 19 Aug 2021 13:35

Tgt Ion: 73 Resp: 10113
 Ion Ratio Lower Upper
 73 100
 43 18.6 14.2 21.4
 57 22.8 17.7 26.5

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKB0DL

Manual Integrations
 APPROVED




MMDadoda
 8/20/2021 12:53:14 PM




#33

Benzene
 Concen: 2.324 ug/L
 RT: 5.108 min Scan# 1265
 Delta R.T. 0.013 min
 Lab File: VV021887.D
 Acq: 19 Aug 2021 13:35

Tgt Ion: 78 Resp: 30683









#42
Toluene
Concen: 11.283 ug/L
RT: 7.391 min Scan# 1975
Delta R.T. 0.003 min
Lab File: VV021887.D
Acq: 19 Aug 2021 13:35

Tgt Ion: 91 Resp: 159082
Ion Ratio Lower Upper
91 100
92 59.2 40.1 74.5

Instrument :
MSVOA_V
ClientSampled :
DBKB0DL




Manual Integrations
APPROVED


MMDadoda
8/20/2021 12:53:14 PM



#52
Ethylbenzene
Concen: 42.900 ug/L
RT: 9.014 min Scan# 2480
Delta R.T. 0.003 min
Lab File: VV021887.D
Acq: 19 Aug 2021 13:35

Tgt Ion: 91 Resp: 645294
Ion Ratio Lower Upper
91 100
106 31.3 22.2 41.2









#53
m,p-Xylene
Concen: 21.750 ug/L
RT: 9.140 min Scan# 2519
Delta R.T. 0.000 min
Lab File: VV021887.D
Acq: 19 Aug 2021 13:35

Tgt Ion:106 Resp: 126793
Ion Ratio Lower Upper
106 100
91 198.5 139.2 258.6

Instrument :
MSVOA_V
ClientSampled :
DBKB0DL




Manual Integrations
APPROVED


MMDadoda
8/20/2021 12:53:14 PM



#54
o-Xylene
Concen: 12.952 ug/L
RT: 9.545 min Scan# 2645
Delta R.T. 0.000 min
Lab File: VV021887.D
Acq: 19 Aug 2021 13:35

Tgt Ion:106 Resp: 74844
Ion Ratio Lower Upper
106 100
91 211.1 148.1 275.0









#55
Styrene
Concen: 9.271 ug/L
RT: 9.564 min Scan# 2651
Delta R.T. 0.003 min
Lab File: VV021887.D
Acq: 19 Aug 2021 13:35

Instrument :
MSVOA_V
ClientSampled :
DBKB0DL

Tgt Ion:104 Resp: 89715
Ion Ratio Lower Upper
104 100
78 43.5 30.2 56.2




Manual Integrations
APPROVED


MMDadoda
8/20/2021 12:53:14 PM



#62
1,3,5-Trimethylbenzene
Concen: 3.115 ug/L
RT: 10.542 min Scan# 2955
Delta R.T. 0.003 min
Lab File: VV021887.D
Acq: 19 Aug 2021 13:35

Tgt Ion:105 Resp: 40604
Ion Ratio Lower Upper
105 100
120 50.3 40.2 60.4





#63
 1,2,4-Trimethylbenzene
 Concen: 9.959 ug/L
 RT: 10.918 min Scan# 3072
 Delta R.T. 0.003 min
 Lab File: VV021887.D
 Acq: 19 Aug 2021 13:35

Tgt Ion:105 Resp: 131979
 Ion Ratio Lower Upper
 105 100
 120 46.4 37.3 55.9

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKB0DL

Manual Integrations
 APPROVED

MMDadoda
 8/20/2021 12:53:14 PM

