

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081921\
 Data File : VW021894.D
 Acq On : 19 Aug 2021 17:12
 Operator : SY/MD
 Sample : M3448-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKG1

Quant Time: Aug 20 02:18:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 20 02:14:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	468190	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	430999	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	203131	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	136774	37.369	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.740%
7) Chloroethane-d5	1.558	69	108085	39.132	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.260%
11) 1,1-Dichloroethene-d2	2.101	63	175480	30.119	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.240%
21) 2-Butanone-d5	3.873	46	204464	89.820	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.820%
24) Chloroform-d	4.349	84	252911	39.883	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.760%
26) 1,2-Dichloroethane-d4	5.034	65	156954	43.029	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.060%
32) Benzene-d6	5.050	84	544583	43.939	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.880%
36) 1,2-Dichloropropane-d6	6.072	67	172263	44.632	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.260%
41) Toluene-d8	7.317	98	494353	43.154	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.300%
43) trans-1,3-Dichloroprop...	7.622	79	74355	39.145	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.280%
47) 2-Hexanone-d5	8.088	63	123584	97.696	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.700%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	222407	46.500	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	191705	45.602	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081921\
Data File : VV021894.D
Acq On : 19 Aug 2021 17:12
Operator : SY/MD
Sample : M3448-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKG1

Quant Time: Aug 20 02:18:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 20 02:14:47 2021
Response via : Initial Calibration

