

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081922\
 Data File : VV027447.D
 Acq On : 19 Aug 2022 17:58
 Operator : SY/MD
 Sample : N4253-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EX9E1

Quant Time: Aug 20 06:06:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 20 05:55:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	128001	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	124433	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	53609	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	43404	4.764	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.200%
7) Chloroethane-d5	1.564	69	40156	6.023	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	120.400%
11) 1,1-Dichloroethene-d2	2.105	63	73738	3.471	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.400%
20) 2-Butanone-d5	3.892	46	69650	61.435	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.880%
24) Chloroform-d	4.346	84	87517	4.805	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
26) 1,2-Dichloroethane-d4	5.031	65	41385	5.252	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.000%
32) Benzene-d6	5.047	84	165633	4.806	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.200%
36) 1,2-Dichloropropane-d6	6.069	67	51182	5.108	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.200%
41) Toluene-d8	7.317	98	131896	4.238	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.800%
43) trans-1,3-Dichloroprop...	7.625	79	15779	4.599	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.000%
46) 2-Hexanone-d5	8.088	63	48154	42.637	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.280%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	30770	4.822	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	41060	5.167	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081922\
Data File : VV027447.D
Acq On : 19 Aug 2022 17:58
Operator : SY/MD
Sample : N4253-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EX9E1

Quant Time: Aug 20 06:06:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 20 05:55:43 2022
Response via : Initial Calibration

