

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV081925\
 Data File : VV038966.D
 Acq On : 19 Aug 2025 15:20
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5326

Quant Time: Aug 20 03:36:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM081925.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 20 03:16:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5238	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5085	0.500	ug/L	0.02
11) 1,4-Dichlorobenzene-d4	11.191	152	2283	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	2726	0.569	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	114.000%
4) 1,2-Dichloroethane-d4	4.969	65	1387	0.493	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
7) 1,2-Dichloropropane-d6	5.995	67	1679	0.599	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	120.000%
8) Toluene-d8	7.253	98	3230	0.514	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1165	0.586	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	118.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.294	62	5150	0.558	ug/L	100
6) Trichloroethene	5.852	95	2554	0.566	ug/L	100
9) 1,2-Dibromoethane	8.305	107	1538	0.571	ug/L	95
12) 1,2,3-Trichloropropane	10.197	75	1285	0.551	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.356	75	288	0.534	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV081925\
Data File : VV038966.D
Acq On : 19 Aug 2025 15:20
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5326

Quant Time: Aug 20 03:36:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM081925.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 20 03:16:00 2025
Response via : Initial Calibration

