

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082018\
 Data File : VV007087.D
 Acq On : 20 Aug 2018 22:13
 Operator : SY/MD
 Sample : J4488-21
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0323

Quant Time: Aug 21 06:51:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 21 05:25:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	241251	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	230550	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	87998	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	76753	4.93	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.60%
7) Chloroethane-d5	1.58	69	71024	5.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.60%
11) 1,1-Dichloroethene-d2	2.14	63	400419	13.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	274.00%#
20) 2-Butanone-d5	3.97	46	231629	51.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.44%
24) Chloroform-d	4.40	84	172133	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.00%
26) 1,2-Dichloroethane-d4	5.09	65	89364	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.00%
32) Benzene-d6	5.10	84	338986	5.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.40%
36) 1,2-Dichloropropane-d6	6.12	67	109396	5.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.00%
41) Toluene-d8	7.37	98	232895	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
43) trans-1,3-Dichloropropene-	7.67	79	22639	3.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.20%
46) 2-Hexanone-d5	8.15	63	87337	30.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	60.64%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	81764	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	83873	5.48	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.60%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Trichlorofluoromethane	1.77	101	10537	0.35	ug/L	98
12) 1,1-Dichloroethene	2.14	96	598289	34.54	ug/L #	70
13) Acetone	2.21	43	12245	5.31	ug/L	94
19) 1,1-Dichloroethane	3.23	63	829611	24.77	ug/L	98
25) Chloroform	4.43	83	12538	0.34	ug/L	97
27) 1,2-Dichloroethane	5.19	62	10928	0.48	ug/L #	85
29) 1,1,1-Trichloroethane	4.66	97	14322928	465.92	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082018\
 Data File : VV007087.D
 Acq On : 20 Aug 2018 22:13
 Operator : SY/MD
 Sample : J4488-21
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 C0323

Quant Time: Aug 21 06:51:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 21 05:25:12 2018
 Response via : Initial Calibration

