

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082018\
 Data File : VV007080.D
 Acq On : 20 Aug 2018 19:07
 Operator : SY/MD
 Sample : J4488-14
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0309

Quant Time: Aug 21 05:54:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 21 05:25:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	236233	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	231436	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	87968	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	80133	5.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.20%
7) Chloroethane-d5	1.58	69	70942	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.60%
11) 1,1-Dichloroethene-d2	2.13	63	111448	3.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.80%
20) 2-Butanone-d5	3.97	46	215524	49.15	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.30%
24) Chloroform-d	4.40	84	172422	5.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.40%
26) 1,2-Dichloroethane-d4	5.09	65	87317	5.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.80%
32) Benzene-d6	5.10	84	321175	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.00%
36) 1,2-Dichloropropane-d6	6.12	67	110441	5.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.60%
41) Toluene-d8	7.37	98	235290	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
43) trans-1,3-Dichloropropene-	7.67	79	21962	3.11	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.20%
46) 2-Hexanone-d5	8.15	63	77206	26.70	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	53.40%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	77807	4.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	83949	5.49	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.80%


Target Compounds					Ovalue
13) Acetone	2.21	43	5512	2.44 ug/L	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082018\
Data File : VV007080.D
Acq On : 20 Aug 2018 19:07
Operator : SY/MD
Sample : J4488-14
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0309

Quant Time: Aug 21 05:54:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 21 05:25:12 2018
Response via : Initial Calibration





#13

Acetone

Concen: 2.44 ug/L

RT: 2.21 min Scan# 410

Delta R.T. 0.00 min

Lab File: VV007080.D

Acq: 20 Aug 2018 19:07

Instrument :
MSVOA_V
ClientSampleId :
C0309

Tot Ion: 43 Resp: 5512

Ion Ratio Lower Upper

43 100

58 29.8 0.0 71.4

