

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082019\
 Data File : VV012310.D
 Acq On : 20 Aug 2019 14:11
 Operator : SY/MD
 Sample : K4373-14DL 10X
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AD2DL

Quant Time: Aug 21 02:19:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR081919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 21 01:13:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	133407	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	122796	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	72352	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	23873	3.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.00%
7) Chloroethane-d5	1.56	69	18093	3.63	ug/L	-0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.60%
11) 1,1-Dichloroethene-d2	2.12	63	42001	3.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.80%
20) 2-Butanone-d5	3.93	46	83210	45.27	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.54%
24) Chloroform-d	4.40	84	83220	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.20%
26) 1,2-Dichloroethane-d4	5.08	65	44067	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
32) Benzene-d6	5.09	84	164321	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
36) 1,2-Dichloropropane-d6	6.11	67	43199	4.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.00%
41) Toluene-d8	7.36	98	163093	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.80%
43) trans-1,3-Dichloropropene-	7.66	79	20361	5.16	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.20%
46) 2-Hexanone-d5	8.13	63	78413	63.00	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	126.00%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	33679	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	72282	4.97	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.40%

Target Compounds

					Ovalue
13) Acetone	2.19	43	1226	1.275	ug/L 78
33) Benzene	5.14	78	172933	5.091	ug/L 100
42) Toluene	7.43	91	321545	8.553	ug/L 99
52) Ethylbenzene	9.05	91	232969	5.538	ug/L 96
53) m,p-xylene	9.18	106	167020	10.171	ug/L 95
54) o-xylene	9.59	106	87627	5.548	ug/L 99
56) Isopropylbenzene	9.97	105	16414	0.377	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082019\
Data File : VV012310.D
Acq On : 20 Aug 2019 14:11
Operator : SY/MD
Sample : K4373-14DL 10X
Misc : 25mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AD2DL

Quant Time: Aug 21 02:19:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR081919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 21 01:13:54 2019
Response via : Initial Calibration

