

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082024\
 Data File : VV036982.D
 Acq On : 20 Aug 2024 11:16
 Operator : SY/MD
 Sample : VV0820WBL01
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK380

Quant Time: Aug 21 02:06:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 21 02:05:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	628464	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	612311	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	289180	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	213837	42.600	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.200%
7) Chloroethane-d5	1.532	69	185366	46.827	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.660%
11) 1,1-Dichloroethene-d2	2.060	65	100117	46.628	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.260%
21) 2-Butanone-d5	3.796	46	224304	84.339	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.340%
24) Chloroform-d	4.246	84	429549	47.686	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.380%
26) 1,2-Dichloroethane-d4	4.937	65	273434	47.212	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.420%
32) Benzene-d6	4.953	84	863014	46.664	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.320%
36) 1,2-Dichloropropane-d6	5.985	67	279974	46.925	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.840%
41) Toluene-d8	7.239	98	768932	45.325	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.660%
43) trans-1,3-Dichloroprop...	7.551	79	121889	39.254	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.500%
47) 2-Hexanone-d5	8.024	63	127796	74.494	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.490%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	309188	45.132	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.260%
66) 1,2-Dichlorobenzene-d4	11.554	152	280314	47.570	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082024\
Data File : VV036982.D
Acq On : 20 Aug 2024 11:16
Operator : SY/MD
Sample : VV0820WBL01
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK380

Quant Time: Aug 21 02:06:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 21 02:05:44 2024
Response via : Initial Calibration

