

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082025\
 Data File : VV038969.D
 Acq On : 20 Aug 2025 10:01
 Operator : SY/MD
 Sample : VV0820WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK231

Quant Time: Aug 21 01:22:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM081925.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 20 03:16:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5798	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	5458	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2433	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	2569	0.484	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	96.000%
4) 1,2-Dichloroethane-d4	4.969	65	1327	0.426	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	86.000%
7) 1,2-Dichloropropane-d6	5.995	67	1565	0.520	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	104.000%
8) Toluene-d8	7.253	98	3019	0.447	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	90.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1080	0.506	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082025\
Data File : VV038969.D
Acq On : 20 Aug 2025 10:01
Operator : SY/MD
Sample : VV0820WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK231

Quant Time: Aug 21 01:22:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM081925.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 20 03:16:00 2025
Response via : Initial Calibration

