

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082118\
 Data File : VV007112.D
 Acq On : 21 Aug 2018 20:09
 Operator : SY/MD
 Sample : J4488-05
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Aug 22 04:15:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR082118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 22 03:46:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	280219	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	276494	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	108188	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	96131	4.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.20%
7) Chloroethane-d5	1.58	69	83727	5.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.40%
11) 1,1-Dichloroethene-d2	2.13	63	132846	3.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.60%
20) 2-Butanone-d5	3.97	46	224354	49.27	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.54%
24) Chloroform-d	4.40	84	193169	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	5.09	65	102254	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.20%
32) Benzene-d6	5.10	84	375361	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.40%
36) 1,2-Dichloropropane-d6	6.12	67	124408	4.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.20%
41) Toluene-d8	7.37	98	282004	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.60%
43) trans-1,3-Dichloropropene-	7.67	79	41218	4.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.40%
46) 2-Hexanone-d5	8.15	63	105934	40.37	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.74%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	93704	5.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	103650	5.03	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082118\
Data File : VV007112.D
Acq On : 21 Aug 2018 20:09
Operator : SY/MD
Sample : J4488-05
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Aug 22 04:15:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR082118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 22 03:46:37 2018
Response via : Initial Calibration

