

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082119\
 Data File : VV012323.D
 Acq On : 21 Aug 2019 19:32
 Operator : SY/MD
 Sample : K4464-09
 Misc : 5mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEJW4

Quant Time: Aug 22 04:02:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 22 03:48:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	218080	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	200564	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	95553	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	49809	48.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.48%
7) Chloroethane-d5	1.58	69	37229	50.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.58%
11) 1,1-Dichloroethene-d2	2.13	63	78445	38.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.18%
21) 2-Butanone-d5	3.93	46	67091	83.37	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.37%
24) Chloroform-d	4.40	84	138151	47.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.16%
26) 1,2-Dichloroethane-d4	5.08	65	103594	51.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.38%
32) Benzene-d6	5.10	84	269270	50.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.68%
36) 1,2-Dichloropropane-d6	6.12	67	73640	50.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.82%
41) Toluene-d8	7.36	98	260642	50.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.00%
43) trans-1,3-Dichloropropene-	7.66	79	44089	49.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.98%
47) 2-Hexanone-d5	8.13	63	54012	90.48	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.48%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	100785	48.93	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.86%
64) 1,2-Dichlorobenzene-d4	11.67	152	108794	51.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.58%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082119\
Data File : VV012323.D
Acq On : 21 Aug 2019 19:32
Operator : SY/MD
Sample : K4464-09
Misc : 5mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEJW4

Quant Time: Aug 22 04:02:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082119WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 22 03:48:59 2019
Response via : Initial Calibration

