

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082119\
 Data File : VV012335.D
 Acq On : 22 Aug 2019 00:25
 Operator : SY/MD
 Sample : K4454-14
 Misc : 5mL/MSVOA V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F2C10

Quant Time: Aug 22 04:57:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 22 03:48:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	217635	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	207613	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	96154	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	48276	46.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.70%
7) Chloroethane-d5	1.58	69	35908	48.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.22%
11) 1,1-Dichloroethene-d2	2.13	63	74583	36.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.54%
21) 2-Butanone-d5	3.93	46	57707	71.85	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	71.85%
24) Chloroform-d	4.40	84	135059	46.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.24%
26) 1,2-Dichloroethane-d4	5.08	65	103853	51.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.84%
32) Benzene-d6	5.10	84	268412	48.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.96%
36) 1,2-Dichloropropane-d6	6.12	67	73235	48.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.82%
41) Toluene-d8	7.36	98	259218	48.04	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.08%
43) trans-1,3-Dichloropropene-	7.66	79	40554	43.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.96%
47) 2-Hexanone-d5	8.13	63	55322	89.53	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.53%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	99810	46.81	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.62%
64) 1,2-Dichlorobenzene-d4	11.67	152	110240	52.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.30%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082119\
Data File : VV012335.D
Acq On : 22 Aug 2019 00:25
Operator : SY/MD
Sample : K4454-14
Misc : 5mL/MSVOA V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F2C10

Quant Time: Aug 22 04:57:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082119WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 22 03:48:59 2019
Response via : Initial Calibration

