

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082119\
 Data File : VV012337.D
 Acq On : 22 Aug 2019 01:14
 Operator : SY/MD
 Sample : K4454-16
 Misc : 5mL/MSVOA V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F2C12

Quant Time: Aug 22 05:05:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 22 03:48:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	215910	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	205631	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	94412	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	46303	45.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.60%
7) Chloroethane-d5	1.58	69	34305	46.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.62%
11) 1,1-Dichloroethene-d2	2.13	63	72092	35.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.66%
21) 2-Butanone-d5	3.93	46	54367	68.23	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	68.23%
24) Chloroform-d	4.40	84	131560	45.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.56%
26) 1,2-Dichloroethane-d4	5.08	65	101348	50.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.16%
32) Benzene-d6	5.10	84	263487	48.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.10%
36) 1,2-Dichloropropane-d6	6.12	67	71769	48.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.80%
41) Toluene-d8	7.36	98	250297	46.83	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.66%
43) trans-1,3-Dichloropropene-	7.66	79	40036	43.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.68%
47) 2-Hexanone-d5	8.13	63	54087	88.38	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.38%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	99130	46.94	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.88%
64) 1,2-Dichlorobenzene-d4	11.67	152	107278	51.69	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.38%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082119\
Data File : VV012337.D
Acq On : 22 Aug 2019 01:14
Operator : SY/MD
Sample : K4454-16
Misc : 5mL/MSVOA V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F2C12

Quant Time: Aug 22 05:05:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082119WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 22 03:48:59 2019
Response via : Initial Calibration

