

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082120\
 Data File : VV018073.D
 Acq On : 21 Aug 2020 10:36
 Operator : SY/MD
 Sample : VV0821MBL01
 Misc : 5.0uL/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK89

Quant Time: Aug 22 00:10:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 22 00:05:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	295057	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	292195	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	143735	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	105620	46.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.84%
7) Chloroethane-d5	1.58	69	89641	48.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.34%
11) 1,1-Dichloroethene-d2	2.12	63	154706	35.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.86%
21) 2-Butanone-d5	3.91	46	124912	84.72	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.72%
24) Chloroform-d	4.37	84	184282	44.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.94%
26) 1,2-Dichloroethane-d4	5.05	65	124722	46.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.72%
32) Benzene-d6	5.07	84	378288	46.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.96%
36) 1,2-Dichloropropane-d6	6.09	67	121868	46.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.58%
41) Toluene-d8	7.33	98	335933	46.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.78%
43) trans-1,3-Dichloropropene-	7.64	79	55558	46.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.06%
47) 2-Hexanone-d5	8.11	63	62303	73.75	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.75%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	139288	40.63	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.26%
64) 1,2-Dichlorobenzene-d4	11.65	152	138865	49.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.18%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082120\
Data File : VV018073.D
Acq On : 21 Aug 2020 10:36
Operator : SY/MD
Sample : VV0821MBL01
Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK89

Quant Time: Aug 22 00:10:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 22 00:05:13 2020
Response via : Initial Calibration

