

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082219\
 Data File : VV012341.D
 Acq On : 22 Aug 2019 12:17
 Operator : SY/MD
 Sample : VV0822WBL01
 Misc : 5mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK67

Quant Time: Aug 23 07:03:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 23 06:55:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	220461	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	215434	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	97689	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	49471	47.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.80%
7) Chloroethane-d5	1.58	69	37384	49.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.92%
11) 1,1-Dichloroethene-d2	2.13	63	68450	33.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.62%
21) 2-Butanone-d5	3.93	46	56030	68.87	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	68.87%
24) Chloroform-d	4.40	84	133780	45.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.18%
26) 1,2-Dichloroethane-d4	5.08	65	101069	49.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.80%
32) Benzene-d6	5.10	84	263473	45.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.72%
36) 1,2-Dichloropropane-d6	6.12	67	77953	50.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.36%
41) Toluene-d8	7.36	98	262097	46.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.62%
43) trans-1,3-Dichloropropene-	7.66	79	43048	44.99	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.98%
47) 2-Hexanone-d5	8.13	63	47576	74.20	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.20%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	103127	46.61	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.22%
64) 1,2-Dichlorobenzene-d4	11.67	152	111251	51.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082219\
Data File : VV012341.D
Acq On : 22 Aug 2019 12:17
Operator : SY/MD
Sample : VV0822WBL01
Misc : 5mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK67

Quant Time: Aug 23 07:03:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082119WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 23 06:55:48 2019
Response via : Initial Calibration

