

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082219\
 Data File : VV012345.D
 Acq On : 22 Aug 2019 14:15
 Operator : SY/MD
 Sample : K4464-10
 Misc : 5mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Aug 23 07:16:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 23 06:55:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	190126	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	178637	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	81733	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	38444	42.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.42%
7) Chloroethane-d5	1.58	69	29801	46.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.36%
11) 1,1-Dichloroethene-d2	2.13	63	63726	35.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.92%
21) 2-Butanone-d5	3.93	46	54527	77.72	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.72%
24) Chloroform-d	4.40	84	121295	47.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.82%
26) 1,2-Dichloroethane-d4	5.08	65	94675	53.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.32%
32) Benzene-d6	5.10	84	231829	48.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.32%
36) 1,2-Dichloropropane-d6	6.12	67	63424	49.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.46%
41) Toluene-d8	7.36	98	226414	48.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.54%
43) trans-1,3-Dichloropropene-	7.66	79	37461	47.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.44%
47) 2-Hexanone-d5	8.13	63	42327	79.61	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.61%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	86377	47.08	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.16%
64) 1,2-Dichlorobenzene-d4	11.67	152	97116	54.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082219\
Data File : VV012345.D
Acq On : 22 Aug 2019 14:15
Operator : SY/MD
Sample : K4464-10
Misc : 5mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Aug 23 07:16:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082119WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 23 06:55:48 2019
Response via : Initial Calibration

