

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082222\
 Data File : VV027468.D
 Acq On : 22 Aug 2022 15:21
 Operator : SY/MD
 Sample : N4252-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EX9C9

Quant Time: Aug 23 03:13:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 23 03:03:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	110637	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	111119	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	48707	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	37377	4.747	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	95.000%	
7) Chloroethane-d5	1.565	69	38589	6.696	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	134.000%#	
11) 1,1-Dichloroethene-d2	2.105	63	69355	3.777	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	75.600%	
20) 2-Butanone-d5	3.889	46	50958	52.002	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	104.000%	
24) Chloroform-d	4.346	84	81976	5.207	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.200%	
26) 1,2-Dichloroethane-d4	5.031	65	37320	5.479	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.600%	
32) Benzene-d6	5.050	84	149499	4.857	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.200%	
36) 1,2-Dichloropropane-d6	6.069	67	46254	5.169	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	103.400%	
41) Toluene-d8	7.317	98	116856	4.205	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.000%	
43) trans-1,3-Dichloroprop...	7.629	79	14333	4.678	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	93.600%	
46) 2-Hexanone-d5	8.092	63	44805	44.425	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	88.840%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	28854	5.064	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	37427	5.184	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.600%	
Target Compounds						Qvalue
3) Chloromethane	1.240	50	4190	0.409	ug/L	91
10) 1,1,2-Trichloro-1,2,2-...	2.114	101	9687	1.047	ug/L	95
18) trans-1,2-Dichloroethene	2.761	96	4218	0.489	ug/L	98
22) cis-1,2-Dichloroethene	3.909	96	121307	13.199	ug/L	94
34) Trichloroethene	5.912	95	107997	11.108	ug/L	96
47) Tetrachloroethene	7.976	164	5472	0.722	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082222\
Data File : VV027468.D
Acq On : 22 Aug 2022 15:21
Operator : SY/MD
Sample : N4252-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EX9C9

Quant Time: Aug 23 03:13:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 23 03:03:07 2022
Response via : Initial Calibration

