

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV082222\  
 Data File : VV027471.D  
 Acq On : 22 Aug 2022 16:35  
 Operator : SY/MD  
 Sample : N4252-11  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 EX9D2

Quant Time: Aug 23 03:13:45 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR081922WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Tue Aug 23 03:03:07 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.616  | 114   | 112832   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.850  | 117   | 112400   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 47622    | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.304  | 65    | 38759    | 4.827    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 96.600%  |
| 7) Chloroethane-d5            | 1.568  | 69    | 37667    | 6.409    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 128.200% |
| 11) 1,1-Dichloroethene-d2     | 2.105  | 63    | 69432    | 3.708    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 74.200%  |
| 20) 2-Butanone-d5             | 3.896  | 46    | 59499    | 59.537   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 119.080% |
| 24) Chloroform-d              | 4.346  | 84    | 82417    | 5.133    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 102.600% |
| 26) 1,2-Dichloroethane-d4     | 5.034  | 65    | 35529    | 5.115    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 102.200% |
| 32) Benzene-d6                | 5.047  | 84    | 146677   | 4.711    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 94.200%  |
| 36) 1,2-Dichloropropane-d6    | 6.066  | 67    | 44981    | 4.969    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 99.400%  |
| 41) Toluene-d8                | 7.317  | 98    | 112850   | 4.014    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 80.200%  |
| 43) trans-1,3-Dichloroprop... | 7.625  | 79    | 13997    | 4.516    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 90.400%  |
| 46) 2-Hexanone-d5             | 8.092  | 63    | 44167    | 43.293   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 86.580%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84    | 28456    | 4.937    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 98.800%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.625 | 152   | 37188    | 5.268    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 105.400% |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV082222\  
Data File : VV027471.D  
Acq On : 22 Aug 2022 16:35  
Operator : SY/MD  
Sample : N4252-11  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
EX9D2

Quant Time: Aug 23 03:13:45 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR081922WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Tue Aug 23 03:03:07 2022  
Response via : Initial Calibration

