

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082222\
 Data File : VV027466.D
 Acq On : 22 Aug 2022 14:32
 Operator : SY/MD
 Sample : N4252-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EX9D6

Quant Time: Aug 23 03:12:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 23 03:03:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	118727	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	116022	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	48939	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	38072	4.506	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.200%
7) Chloroethane-d5	1.565	69	40337	6.523	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	130.400%#
11) 1,1-Dichloroethene-d2	2.105	63	73242	3.717	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.400%
20) 2-Butanone-d5	3.899	46	39649	37.704	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	75.400%
24) Chloroform-d	4.346	84	84623	5.009	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	5.031	65	38184	5.224	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.047	84	153264	4.769	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
36) 1,2-Dichloropropane-d6	6.069	67	47036	5.034	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.600%
41) Toluene-d8	7.317	98	120019	4.136	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.800%
43) trans-1,3-Dichloroprop...	7.625	79	14559	4.551	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.000%
46) 2-Hexanone-d5	8.092	63	45437	43.148	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.300%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	28773	4.836	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	37416	5.157	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082222\
Data File : VV027466.D
Acq On : 22 Aug 2022 14:32
Operator : SY/MD
Sample : N4252-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EX9D6

Quant Time: Aug 23 03:12:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 23 03:03:07 2022
Response via : Initial Calibration

