

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082225\
 Data File : VV039013.D
 Acq On : 22 Aug 2025 13:26
 Operator : SY/MD
 Sample : Q2929-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DA5-SW4

Quant Time: Aug 25 02:21:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 04:15:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	582739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1134821	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	1033198	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	449917	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	486168	59.457	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	118.920%#
35) Dibromofluoromethane	4.503	113	386669	45.434	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	90.860%
50) Toluene-d8	7.236	98	1362042	45.896	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	91.800%
62) 4-Bromofluorobenzene	10.001	95	501660	46.208	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	92.420%
Target Compounds						
16) Acetone	2.124	43	45975	14.009	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082225\
Data File : VV039013.D
Acq On : 22 Aug 2025 13:26
Operator : SY/MD
Sample : Q2929-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DA5-SW4

Quant Time: Aug 25 02:21:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 04:15:04 2025
Response via : Initial Calibration

