

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV082318\
 Data File : VV007147.D
 Acq On : 23 Aug 2018 18:29
 Operator : SY/MD
 Sample : J4599-13
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETOK9

Quant Time: Aug 24 02:11:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 24 02:07:07 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	515166	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	478415	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	172585	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	144640	44.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.36%
7) Chloroethane-d5	1.57	69	124842	59.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	119.92%
11) 1,1-Dichloroethene-d2	2.13	63	208035	34.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.40%
21) 2-Butanone-d5	3.96	46	271399	98.50	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.50%
24) Chloroform-d	4.41	84	331029	47.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.58%
26) 1,2-Dichloroethane-d4	5.09	65	208382	49.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.24%
32) Benzene-d6	5.10	84	656032	49.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.68%
36) 1,2-Dichloropropane-d6	6.12	67	222120	50.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.00%
41) Toluene-d8	7.37	98	507627	43.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.78%
43) trans-1,3-Dichloropropene-	7.67	79	81299	43.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.34%
47) 2-Hexanone-d5	8.14	63	114460	73.18	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.18%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	284943	44.53	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.06%
64) 1,2-Dichlorobenzene-d4	11.68	152	178526	51.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.90%

Target Compounds


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Vinyl chloride	1.32	62	22990	4.699	ug/L	87
13) Acetone	2.20	43	8036	3.718	ug/L	97
19) 1,1-Dichloroethane	3.23	63	49716	6.495	ug/L	98
20) cis-1,2-Dichloroethene	3.97	96	39732	9.193	ug/L	96
48) 2-Hexanone	8.19	43	41731	9.705	ug/L	89
59) 1,2,3-Trichloropropane	11.30	75	19173	3.393	ug/L	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV082318\
Data File : VV007147.D
Acq On : 23 Aug 2018 18:29
Operator : SY/MD
Sample : J4599-13
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETOK9

Quant Time: Aug 24 02:11:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082218WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 24 02:07:07 2018
Response via : Initial Calibration





#5

Vinyl chloride

Concen: 4.699 ug/L

RT: 1.32 min Scan# 134

Delta R.T. -0.00 min

Lab File: VV007147.D

Acq: 23 Aug 2018 18:29

Instrument :

MSVOA_V

ClientSampleId :





ETOK9

Tgt Ion: 62 Resp: 22990

Ion Ratio Lower Upper

62 100

64 40.8 23.4 43.4



#13

Acetone

Concen: 3.718 ug/L

RT: 2.20 min Scan# 407

Delta R.T. 0.00 min

Lab File: VV007147.D




Acq: 23 Aug 2018 18:29


Tgt Ion: 43 Resp: 8036

Ion Ratio Lower Upper

43 100

58 34.7 0.0 73.0





#19

1,1-Dichloroethane

Concen: 6.495 ug/L

RT: 3.23 min Scan# 728

Delta R.T. 0.00 min

Lab File: VV007147.D

Acq: 23 Aug 2018 18:29

Instrument :

MSVOA_V

ClientSampleId :

ETOK9


Tgt Ion: 63 Resp: 49716

Ion Ratio Lower Upper

63 100

65 33.5 22.7 42.3




83 13.9 9.4 17.6



Abundance Ion 63.00 (62.70 to 63.70): VV007147.D (-634) (-)

Ion 65.00 (64.70 to 65.70): VV007147.D (-634) (-)

Ion 83.00 (82.70 to 83.70): VV007147.D (-634) (-)



#20

cis-1,2-Dichloroethene

Concen: 9.193 ug/L

RT: 3.97 min Scan# 957

Delta R.T. 0.00 min

Lab File: VV007147.D

Acq: 23 Aug 2018 18:29


Tgt Ion: 96 Resp: 39732

Ion Ratio Lower Upper

96 100

61 122.9 89.0 165.4



68 0.0 0.0 0.0




Abundance Ion 96.00 (95.70 to 96.70): VV007147.D (-863) (-)

Ion 61.00 (60.70 to 61.70): VV007147.D (-863) (-)

Ion 68.00 (67.70 to 68.70): VV007147.D (-863) (-)









#48
 2-Hexanone
 Concen: 9.705 ug/L
 RT: 8.19 min Scan# 2271
 Delta R.T. -0.00 min
 Lab File: VV007147.D
 Acq: 23 Aug 2018 18:29

Instrument :
 MSVOA_V
 ClientSampleId :
 ETOK9

Tgt Ion:	43	Resp:	41731
Ion	Ratio	Lower	Upper
43	100		
58	50.0	47.9	71.9
57	16.0	15.8	23.8
100	11.0	10.6	16.0



#59
 1,2,3-Trichloropropane
 Concen: 3.393 ug/L
 RT: 11.30 min Scan# 3237
 Delta R.T. 0.97 min
 Lab File: VV007147.D
 Acq: 23 Aug 2018 18:29

Tgt Ion:	75	Resp:	19173
Ion	Ratio	Lower	Upper
75	100		
77	38.2	25.8	38.8

