

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082318\
 Data File : VV007141.D
 Acq On : 23 Aug 2018 15:38
 Operator : SY/MD
 Sample : J4599-08
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETOL2

Manual Integrations
 APPROVED

sam
 8/24/2018 3:00:45 PM

Quant Time: Aug 24 02:00:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 24 01:26:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	481105	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	461239	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	158053	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	141597	46.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.66%
7) Chloroethane-d5	1.56	69	116553m	59.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	119.88%
11) 1,1-Dichloroethene-d2	2.13	63	197166	35.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.44%
21) 2-Butanone-d5	3.95	46	268895	104.50	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.50%
24) Chloroform-d	4.40	84	311563	48.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.34%
26) 1,2-Dichloroethane-d4	5.09	65	198149	50.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.06%
32) Benzene-d6	5.10	84	620016	48.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.74%
36) 1,2-Dichloropropane-d6	6.12	67	212253	50.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.10%
41) Toluene-d8	7.36	98	484766	42.98	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.96%
43) trans-1,3-Dichloropropene-	7.67	79	81690	45.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.00%
47) 2-Hexanone-d5	8.14	63	114889	76.19	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.19%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	288226	46.72	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.44%
64) 1,2-Dichlorobenzene-d4	11.68	152	166971	52.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.08%

Target Compounds					Ovalue
13) Acetone	2.19	43	7393	3.66 ug/L	99
20) cis-1,2-Dichloroethene	3.96	96	5624	1.39 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082318\
Data File : VV007141.D
Acq On : 23 Aug 2018 15:38
Operator : SY/MD
Sample : J4599-08
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETOL2

Manual Integrations
APPROVED

sam
8/24/2018 3:00:45 PM

Quant Time: Aug 24 02:00:42 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082218WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 24 01:26:08 2018
Response via : Initial Calibration

