

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082318\
 Data File : VV007155.D
 Acq On : 23 Aug 2018 21:48
 Operator : SY/MD
 Sample : J4622-11
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETBK3

Quant Time: Aug 24 02:45:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 24 02:07:07 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	541958	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	512276	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	188911	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	142645	41.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.76%
7) Chloroethane-d5	1.57	69	123320	56.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	112.60%
11) 1,1-Dichloroethene-d2	2.13	63	205884	32.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.30%
21) 2-Butanone-d5	3.95	46	274435	94.67	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.67%
24) Chloroform-d	4.41	84	341532	46.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.74%
26) 1,2-Dichloroethane-d4	5.09	65	218960	49.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.12%
32) Benzene-d6	5.10	84	666165	46.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.58%
36) 1,2-Dichloropropane-d6	6.12	67	227123	48.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.46%
41) Toluene-d8	7.37	98	512810	40.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.88%
43) trans-1,3-Dichloropropene-	7.67	79	83457	41.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.78%
47) 2-Hexanone-d5	8.14	63	120939	72.22	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	72.22%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	310828	45.36	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.72%
64) 1,2-Dichlorobenzene-d4	11.68	152	192393	50.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.30%

Target Compounds					Ovalue
13) Acetone	2.20	43	8337	3.67 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082318\
Data File : VV007155.D
Acq On : 23 Aug 2018 21:48
Operator : SY/MD
Sample : J4622-11
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETBK3

Quant Time: Aug 24 02:45:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM082218WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 24 02:07:07 2018
Response via : Initial Calibration

