

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082319\
 Data File : VV012377.D
 Acq On : 23 Aug 2019 21:23
 Operator : SY/MD
 Sample : K4500-03
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GANJ9

Quant Time: Aug 26 10:00:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR082319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Aug 26 07:59:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	246058	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	239200	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	95671	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	66612	5.42	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.40%
7) Chloroethane-d5	1.58	69	47516	5.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.80%
11) 1,1-Dichloroethene-d2	2.13	63	87859	4.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.40%
20) 2-Butanone-d5	3.93	46	170654	51.49	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.98%
24) Chloroform-d	4.40	84	142260	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
26) 1,2-Dichloroethane-d4	5.08	65	68322	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
32) Benzene-d6	5.10	84	285014	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	6.12	67	86474	4.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.00%
41) Toluene-d8	7.36	98	233646	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
43) trans-1,3-Dichloropropene-	7.67	79	28183	4.22	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.40%
46) 2-Hexanone-d5	8.13	63	92708	44.68	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.36%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	53710	4.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	82501	5.18	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.60%

Target Compounds

					Ovalue
13) Acetone	2.19	43	11874	8.281 ug/L	97
18) trans-1,2-Dichloroethene	2.79	96	6802	0.463 ug/L	92
22) cis-1,2-Dichloroethene	3.96	96	102765	6.419 ug/L #	99
34) Trichloroethene	5.96	95	23683	1.439 ug/L	94
47) Tetrachloroethene	8.02	164	18040	1.358 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082319\
Data File : VV012377.D
Acq On : 23 Aug 2019 21:23
Operator : SY/MD
Sample : K4500-03
Misc : 25mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GANJ9

Quant Time: Aug 26 10:00:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR082319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Aug 26 07:59:52 2019
Response via : Initial Calibration

