

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW082323\
 Data File : VW031873.D
 Acq On : 23 Aug 2023 16:22
 Operator : SY/MD
 Sample : 04037-09ME
 Misc : 5.13g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EZYJ3ME

Quant Time: Aug 24 01:58:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 24 01:40:47 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Romaben Patel 08/24/2023
 Supervised By :Mahesh Dadoda 08/24/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	295826	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	287336	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	148202	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	69353	37.460	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	74.920%		
7) Chloroethane-d5	1.529	69	45121	27.802	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	55.600%#		
11) 1,1-Dichloroethene-d2	2.040	65	32478	37.464	ug/L	-0.02
Spiked Amount 50.000	Range 60 - 125		Recovery =	74.920%		
21) 2-Butanone-d5	3.818	46	110338	75.776	ug/L	0.03
Spiked Amount 100.000	Range 40 - 130		Recovery =	75.780%		
24) Chloroform-d	4.252	84	161666	39.951	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	79.900%		
26) 1,2-Dichloroethane-d4	4.944	65	111967	41.693	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	83.380%		
32) Benzene-d6	4.960	84	337726	43.074	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.140%		
36) 1,2-Dichloropropane-d6	5.992	67	111354	44.125	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	88.240%		
41) Toluene-d8	7.246	98	306472	42.422	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	84.840%		
43) trans-1,3-Dichloroprop...	7.558	79	59184	45.247	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	90.500%		
47) 2-Hexanone-d5	8.072	63	61371m	79.344	ug/L	0.05
Spiked Amount 100.000	Range 45 - 130		Recovery =	79.340%		
56) 1,1,2,2-Tetrachloroeth...	10.165	84	138759	42.550	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	85.100%		
66) 1,2-Dichlorobenzene-d4	11.570	152	119254	44.141	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	88.280%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082323\
Data File : VV031873.D
Acq On : 23 Aug 2023 16:22
Operator : SY/MD
Sample : 04037-09ME
Misc : 5.13g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EZYJ3ME

Quant Time: Aug 24 01:58:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 24 01:40:47 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/24/2023
Supervised By :Mahesh Dadoda 08/24/2023

