

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW082422\  
 Data File : VW027491.D  
 Acq On : 24 Aug 2022 13:26  
 Operator : SY/MD  
 Sample : N4268-03  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VHBLK001

Quant Time: Aug 25 06:26:24 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR081922WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Aug 25 06:25:30 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	122393	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	121603	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	54089	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	38438	4.413	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery =	88.200%		
7) Chloroethane-d5	1.564	69	38315	6.010	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery =	120.200%		
11) 1,1-Dichloroethene-d2	2.105	63	71466	3.518	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery =	70.400%		
20) 2-Butanone-d5	3.895	46	35662	32.897	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery =	65.800%		
24) Chloroform-d	4.342	84	78700	4.518	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery =	90.400%		
26) 1,2-Dichloroethane-d4	5.027	65	33890	4.498	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery =	90.000%		
32) Benzene-d6	5.043	84	140619	4.175	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery =	83.400%		
36) 1,2-Dichloropropane-d6	6.066	67	42830	4.374	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery =	87.400%		
41) Toluene-d8	7.310	98	112668	3.705	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery =	74.000%		
43) trans-1,3-Dichloroprop...	7.622	79	13656	4.073	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery =	81.400%		
46) 2-Hexanone-d5	8.088	63	42432	38.445	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery =	76.880%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	25357	4.066	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery =	81.400%		
66) 1,2-Dichlorobenzene-d4	11.622	152	39984	4.987	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery =	99.800%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV082422\  
 Data File : VV027491.D  
 Acq On : 24 Aug 2022 13:26  
 Operator : SY/MD  
 Sample : N4268-03  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VHBLK001

Quant Time: Aug 25 06:26:24 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR081922WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Aug 25 06:25:30 2022  
 Response via : Initial Calibration

