

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082423\
 Data File : VV031906.D
 Acq On : 24 Aug 2023 14:02
 Operator : SY/MD
 Sample : VV0824WBL03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK357

Quant Time: Aug 25 00:57:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR082323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 25 00:56:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	122919	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	115209	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	60952	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	19402	4.673	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.400%
7) Chloroethane-d5	1.535	69	23248	4.888	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.800%
11) 1,1-Dichloroethene-d2	2.063	65	12193	4.990	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.800%
20) 2-Butanone-d5	3.815	46	84756	42.174	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.340%
24) Chloroform-d	4.252	84	65685	4.675	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%
26) 1,2-Dichloroethane-d4	4.947	65	34482	4.763	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
32) Benzene-d6	4.963	84	130284	4.970	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.400%
36) 1,2-Dichloropropane-d6	5.992	67	44205	4.782	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.600%
41) Toluene-d8	7.246	98	117206	4.818	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
43) trans-1,3-Dichloroprop...	7.561	79	15722	4.664	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.200%
46) 2-Hexanone-d5	8.030	63	82687	46.006	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.020%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	30052	4.531	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	11.564	152	45801	4.640	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.800%
Target Compounds						
16) Methylene chloride	2.452	84	2330	0.259	ug/L	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082423\
Data File : VV031906.D
Acq On : 24 Aug 2023 14:02
Operator : SY/MD
Sample : VV0824WBL03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK357

Quant Time: Aug 25 00:57:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR082323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 25 00:56:34 2023
Response via : Initial Calibration

