

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082521\
 Data File : VV021941.D
 Acq On : 25 Aug 2021 10:49
 Operator : SY/MD
 Sample : VV0825MBL01
 Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK238

Quant Time: Aug 26 05:18:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 26 05:17:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	378822	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	357641	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	181399	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	112515	37.993	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.980%
7) Chloroethane-d5	1.564	69	95240	42.616	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.240%
11) 1,1-Dichloroethene-d2	2.104	63	138949	29.475	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.960%#
21) 2-Butanone-d5	3.876	46	183856	99.821	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.820%
24) Chloroform-d	4.352	84	227423	44.324	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.640%
26) 1,2-Dichloroethane-d4	5.034	65	142950	48.435	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.880%
32) Benzene-d6	5.053	84	477560	46.435	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.880%
36) 1,2-Dichloropropane-d6	6.072	67	152859	47.728	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.460%
41) Toluene-d8	7.320	98	444349	46.745	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.500%
43) trans-1,3-Dichloroprop...	7.625	79	69480	44.081	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.160%
47) 2-Hexanone-d5	8.088	63	113310	107.947	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.950%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	187284	47.189	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.380%
66) 1,2-Dichlorobenzene-d4	11.628	152	185318	49.364	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082521\
Data File : VV021941.D
Acq On : 25 Aug 2021 10:49
Operator : SY/MD
Sample : VV0825MBL01
Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK238

Quant Time: Aug 26 05:18:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 26 05:17:05 2021
Response via : Initial Calibration

