

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082522\
 Data File : VV027496.D
 Acq On : 25 Aug 2022 15:46
 Operator : SY/MD
 Sample : N4370-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5F15

Quant Time: Aug 26 00:59:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 26 00:58:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	128834	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	130065	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	57945	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	31401	3.425	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	68.400%	
7) Chloroethane-d5	1.564	69	35949	5.357	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	107.200%	
11) 1,1-Dichloroethene-d2	2.105	63	64693	3.026	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	60.600%	
20) 2-Butanone-d5	3.892	46	63978	56.067	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	112.140%	
24) Chloroform-d	4.339	84	74314	4.053	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.000%	
26) 1,2-Dichloroethane-d4	5.027	65	35010	4.414	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.200%	
32) Benzene-d6	5.043	84	131540	3.651	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	73.000%	
36) 1,2-Dichloropropane-d6	6.063	67	41882	3.999	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	80.000%	
41) Toluene-d8	7.310	98	106873	3.285	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	65.800%#	
43) trans-1,3-Dichloroprop...	7.622	79	12851	3.583	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	71.600%	
46) 2-Hexanone-d5	8.088	63	45532	38.570	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	77.140%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	27554	4.131	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	82.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	35614	4.146	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	83.000%	
Target Compounds						
13) Acetone	2.179	43	4228	3.880	ug/L	90
14) Carbon disulfide	2.291	76	2828	0.106	ug/L #	93
18) trans-1,2-Dichloroethene	2.757	96	685	0.068	ug/L	85
22) cis-1,2-Dichloroethene	3.905	96	9908	0.926	ug/L	97
34) Trichloroethene	5.912	95	19803	1.740	ug/L	88
47) Tetrachloroethene	7.973	164	3198	0.361	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082522\
Data File : VV027496.D
Acq On : 25 Aug 2022 15:46
Operator : SY/MD
Sample : N4370-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5F15

Quant Time: Aug 26 00:59:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 26 00:58:54 2022
Response via : Initial Calibration

