

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW082522\
 Data File : VW027524.D
 Acq On : 26 Aug 2022 03:39
 Operator : SY/MD
 Sample : N4370-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5F13

Quant Time: Aug 26 04:01:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 26 01:50:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	125522	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	126111	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56859	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	33683	3.770	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.400%
7) Chloroethane-d5	1.568	69	34307	5.247	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.000%
11) 1,1-Dichloroethene-d2	2.108	63	61174	2.937	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.800%#
20) 2-Butanone-d5	3.899	46	38007	34.186	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	68.380%
24) Chloroform-d	4.346	84	74611	4.177	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.600%
26) 1,2-Dichloroethane-d4	5.034	65	33831	4.378	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
32) Benzene-d6	5.050	84	127847	3.660	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.200%
36) 1,2-Dichloropropane-d6	6.066	67	40217	3.960	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.200%
41) Toluene-d8	7.317	98	105577	3.347	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	67.000%#
43) trans-1,3-Dichloroprop...	7.629	79	10700	3.077	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.600%
46) 2-Hexanone-d5	8.092	63	44072	38.503	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.000%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	27055	4.183	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	35814	4.249	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082522\
Data File : VV027524.D
Acq On : 26 Aug 2022 03:39
Operator : SY/MD
Sample : N4370-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5F13

Quant Time: Aug 26 04:01:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 26 01:50:00 2022
Response via : Initial Calibration

