




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082522\
 Data File : VV027502.D
 Acq On : 25 Aug 2022 18:14
 Operator : SY/MD
 Sample : N4370-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5E85

Manual IntegrationsAPPROVED

Quant Time: Aug 26 01:01:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 26 00:58:54 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 08/27/2022
 Supervised By :Semsettin Yesilyurt 08/27/2022



TIC: VV027502.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

2.185min (+ 0.077) 0.20 ug/L

response 3828

Ion	Exp%	Act%
63.00	100.00	100.00
97.95	70.00	60.95
65.10	22.40	22.49
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082522\
Data File : VV027502.D
Acq On : 25 Aug 2022 18:14
Operator : SY/MD
Sample : N4370-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5E85

Manual IntegrationsAPPROVED

Quant Time: Aug 26 01:01:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 26 00:58:54 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 08/27/2022
Supervised By :Semsettin Yesilyurt 08/27/2022



TIC: VV027502.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

2.105min (-0.003) 3.04 ug/L m

response 58467

Ion	Exp%	Act%
63.00	100.00	100.00
97.95	70.00	3.99#
65.10	22.40	1.47#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082522\
 Data File : VV027502.D
 Acq On : 25 Aug 2022 18:14
 Operator : SY/MD
 Sample : N4370-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5E85

Manual IntegrationsAPPROVED

Quant Time: Aug 26 01:01:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 26 00:58:54 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 08/27/2022
 Supervised By :Semsettin Yesilyurt 08/27/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	115898	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	102624	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	45324	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	34357	4.165	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	83.400%		
7) Chloroethane-d5	1.564	69	29703	4.920	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	98.400%		
11) 1,1-Dichloroethene-d2	2.105	63	58467m	3.040	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	60.800%		
20) 2-Butanone-d5	3.902	46	36187	35.252	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery =	70.500%		
24) Chloroform-d	4.346	84	67316	4.081	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	81.600%		
26) 1,2-Dichloroethane-d4	5.034	65	31341	4.392	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	87.800%		
32) Benzene-d6	5.047	84	116664	4.104	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	82.000%		
36) 1,2-Dichloropropane-d6	6.069	67	39633	4.796	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	96.000%		
41) Toluene-d8	7.317	98	97478	3.798	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	76.000%		
43) trans-1,3-Dichloroprop...	7.628	79	11805	4.172	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	83.400%		
46) 2-Hexanone-d5	8.091	63	40925	43.937	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	87.880%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	25789	4.900	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	98.000%		
66) 1,2-Dichlorobenzene-d4	11.625	152	32655	4.860	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	97.200%		
Target Compounds						
						Qvalue
17) Methyl tert-butyl Ether	2.767	73	1686231	92.941	ug/L	99
34) Trichloroethene	5.924	95	3904	0.435	ug/L	99
47) Tetrachloroethene	7.976	164	6993	1.000	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082522\
Data File : VV027502.D
Acq On : 25 Aug 2022 18:14
Operator : SY/MD
Sample : N4370-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5E85

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 08/27/2022
Supervised By :Semsettin Yesilyurt 08/27/2022

Quant Time: Aug 26 01:01:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 26 00:58:54 2022
Response via : Initial Calibration

