

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082523\
 Data File : VV031925.D
 Acq On : 25 Aug 2023 10:37
 Operator : SY/MD
 Sample : VV0825WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK382

Quant Time: Aug 26 00:50:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 26 00:50:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	250484	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	239118	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	125930	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	63256	40.352	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.700%
7) Chloroethane-d5	1.535	69	59824	43.534	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.060%
11) 1,1-Dichloroethene-d2	2.063	65	31411	42.792	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.580%
21) 2-Butanone-d5	3.799	46	122985	99.750	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.750%
24) Chloroform-d	4.252	84	151523	44.222	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.440%
26) 1,2-Dichloroethane-d4	4.947	65	100622	44.251	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.500%
32) Benzene-d6	4.963	84	310531	47.592	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.180%
36) 1,2-Dichloropropane-d6	5.992	67	104154	49.594	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.180%
41) Toluene-d8	7.246	98	277987	46.239	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.480%
43) trans-1,3-Dichloroprop...	7.558	79	48552	44.604	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.200%
47) 2-Hexanone-d5	8.030	63	61324	95.271	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.270%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	134857	49.692	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.380%
66) 1,2-Dichlorobenzene-d4	11.564	152	108685	47.344	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082523\
Data File : VV031925.D
Acq On : 25 Aug 2023 10:37
Operator : SY/MD
Sample : VV0825WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK382

Quant Time: Aug 26 00:50:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 26 00:50:15 2023
Response via : Initial Calibration

