


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082523\
Data File : VV031922.D
Acq On : 25 Aug 2023 08:52
Operator : SY/MD
Sample : BFB206
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB206

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Title : VOC Analysis
Last Update : Sat Aug 26 00:50:15 2023



AutoFind: Scans 2790, 2791, 2792; Background Corrected with Scan 2779

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.0	8868	PASS
75	95	30	80	53.3	23603	PASS
95	95	100	100	100.0	44315	PASS
96	95	5	9	6.8	2994	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	72.3	32029	PASS
175	174	5	9	8.4	2688	PASS
176	174	95	105	99.2	31781	PASS
177	176	5	10	6.3	2003	PASS

m/z	Abundance
35.90	482.0
37.00	2276.0
38.00	2142.0
39.00	874.0
40.00	175.0
41.00	145.0
42.20	144.0
42.80	134.0
44.00	847.0
45.00	469.0
47.00	673.0
47.90	414.0
49.00	2049.0
50.00	9176.0
51.00	2909.0
55.10	225.0
56.00	853.0
57.00	1436.0
58.10	103.0
60.00	583.0
61.00	2363.0
62.10	2175.0
63.00	1680.0
63.90	157.0
67.00	177.0
68.00	4657.0
69.00	4932.0
70.00	547.0
71.90	254.0
73.00	2353.0
74.00	8142.0
75.10	24424.0
76.00	2181.0
77.00	354.0
78.00	288.0
78.90	1404.0
79.90	444.0
80.90	1445.0
81.90	394.0
87.00	2087.0
88.00	1957.0
91.00	192.0
92.00	1381.0
93.00	1880.0
94.00	5648.0
95.00	45752.0
96.10	3077.0
96.90	148.0
103.90	148.0
104.90	102.0
105.90	245.0
115.90	128.0
116.90	280.0
117.80	179.0
118.90	254.0
128.00	215.0
129.80	169.0
140.90	437.0
143.00	464.0
171.80	101.0
174.00	33072.0
175.00	2796.0
176.00	32912.0
177.00	2078.0

Instrument :
MSVOA_V
ClientSampleId :
BFB206

m/z	Abundance
35.90	479.0
37.00	2122.0
38.00	1914.0
39.00	811.0
40.00	192.0
41.00	124.0
42.20	130.0
42.80	143.0
44.00	831.0
45.00	461.0
47.00	601.0
47.90	344.0
49.00	1852.0
50.00	8423.0
51.00	2704.0
55.10	244.0
56.00	773.0
57.00	1370.0
58.10	102.0
60.00	548.0
61.00	2184.0
62.10	1987.0
63.00	1570.0
63.90	143.0
68.00	4434.0
69.00	4721.0
70.00	517.0
71.90	272.0
73.00	2310.0
74.00	7562.0
75.10	22952.0
76.10	1973.0
77.00	379.0
78.00	261.0
78.90	1327.0
79.90	440.0
80.90	1320.0
81.90	368.0
87.00	1984.0
88.00	1758.0
91.00	181.0
92.00	1239.0
93.00	1739.0
94.00	5177.0
95.00	43384.0
96.10	2746.0
96.90	116.0
104.00	148.0
105.00	106.0
105.90	224.0
115.70	114.0
116.10	116.0
116.90	293.0
117.90	177.0
118.90	235.0
127.90	214.0
129.90	163.0
134.80	100.0
140.90	412.0
143.00	451.0
148.00	109.0
156.90	108.0
171.70	100.0
174.00	32112.0
175.00	2623.0
176.00	31768.0
177.00	1930.0

Instrument :
MSVOA_V
ClientSampleId :
BFB206

m/z	Abundance
39.00	122.0
40.00	178.0
40.90	104.0
42.00	108.0
42.80	102.0
44.00	515.0
55.00	121.0
56.90	137.0
67.00	100.0
77.00	115.0

Instrument :
MSVOA_V
ClientSampleId :
BFB206

m/z	Abundance
36.00	430.0
37.00	2274.0
38.00	2216.0
39.00	868.0
39.90	160.0
41.00	181.0
42.20	130.0
42.80	133.0
44.00	840.0
45.00	408.0
47.00	687.0
47.90	398.0
49.00	1969.0
50.00	9006.0
51.00	2876.0
51.90	104.0
55.00	197.0
56.00	845.0
57.00	1388.0
60.00	560.0
61.00	2341.0
62.00	2296.0
63.00	1627.0
63.90	153.0
67.00	209.0
68.00	4662.0
69.10	4681.0
70.00	485.0
72.00	245.0
73.00	2171.0
74.00	8024.0
75.10	23432.0
76.00	2100.0
77.00	336.0
78.00	260.0
78.90	1351.0
79.90	448.0
80.90	1409.0
81.90	385.0
87.00	1992.0
88.00	2045.0
91.00	207.0
92.00	1431.0
93.00	1905.0
94.00	5452.0
95.00	43808.0
96.10	3159.0
97.00	134.0
103.90	137.0
104.90	101.0
105.90	222.0
115.90	137.0
116.80	257.0
117.90	170.0
118.90	247.0
127.90	191.0
129.80	175.0
140.90	414.0
143.00	422.0
174.00	30904.0
175.00	2645.0
176.00	30664.0
177.00	2002.0

Instrument :
MSVOA_V
ClientSampleId :
BFB206