

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082525\
 Data File : VV039027.D
 Acq On : 25 Aug 2025 12:11
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Aug 26 03:22:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082525W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 26 02:37:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	422801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	880620	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	795549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	349497	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	365186	59.885	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	119.760%#
35) Dibromofluoromethane	4.506	113	314439	49.582	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	99.160%
50) Toluene-d8	7.236	98	1084138	48.281	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	96.560%
62) 4-Bromofluorobenzene	10.001	95	355190	43.170	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	86.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082525\
Data File : VV039027.D
Acq On : 25 Aug 2025 12:11
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Aug 26 03:22:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082525W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 26 02:37:06 2025
Response via : Initial Calibration

