

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082621\
 Data File : VV021957.D
 Acq On : 26 Aug 2021 11:37
 Operator : SY/MD
 Sample : VV0826WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK240

Quant Time: Aug 27 03:48:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 27 03:46:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	394951	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	374049	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	199954	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	114539	37.097	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.200%
7) Chloroethane-d5	1.564	69	98045	42.080	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.160%
11) 1,1-Dichloroethene-d2	2.104	63	156077	31.757	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.520%
21) 2-Butanone-d5	3.870	46	197083	102.632	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.630%
24) Chloroform-d	4.346	84	239930	44.852	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.700%
26) 1,2-Dichloroethane-d4	5.030	65	152512	49.565	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.120%
32) Benzene-d6	5.047	84	524143	48.729	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.460%
36) 1,2-Dichloropropane-d6	6.069	67	166890	49.824	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.640%
41) Toluene-d8	7.317	98	488995	49.185	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.380%
43) trans-1,3-Dichloroprop...	7.622	79	75165	45.596	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.200%
47) 2-Hexanone-d5	8.088	63	119235	108.609	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.610%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	199921	48.163	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.320%
66) 1,2-Dichlorobenzene-d4	11.625	152	204692	49.465	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.940%
Target Compounds						
					Qvalue	
25) Chloroform	4.378	83	12592	2.400	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082621\
Data File : VV021957.D
Acq On : 26 Aug 2021 11:37
Operator : SY/MD
Sample : VV0826WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK240

Quant Time: Aug 27 03:48:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 27 03:46:52 2021
Response via : Initial Calibration

