

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW082622\
 Data File : VW027529.D
 Acq On : 26 Aug 2022 12:44
 Operator : SY/MD
 Sample : N4367-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5F26

Quant Time: Aug 27 03:13:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 27 03:12:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	133283	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	129521	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56508	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	33423	3.523	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	70.400%	
7) Chloroethane-d5	1.564	69	36370	5.239	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.800%	
11) 1,1-Dichloroethene-d2	2.105	63	64017	2.894	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	57.800%#	
20) 2-Butanone-d5	3.895	46	45825	38.818	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	77.640%	
24) Chloroform-d	4.342	84	78827	4.156	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.200%	
26) 1,2-Dichloroethane-d4	5.030	65	37759	4.602	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.000%	
32) Benzene-d6	5.047	84	139787	3.896	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	78.000%	
36) 1,2-Dichloropropane-d6	6.066	67	42944	4.117	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	82.400%	
41) Toluene-d8	7.313	98	112909	3.486	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	69.800%#	
43) trans-1,3-Dichloroprop...	7.628	79	13348	3.738	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.800%	
46) 2-Hexanone-d5	8.091	63	48959	41.647	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	83.300%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	27596	4.155	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	83.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	38448	4.590	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	91.800%	
Target Compounds						Qvalue
34) Trichloroethene	5.921	95	3547	0.313	ug/L	88
47) Tetrachloroethene	7.972	164	50397	5.708	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082622\
Data File : VV027529.D
Acq On : 26 Aug 2022 12:44
Operator : SY/MD
Sample : N4367-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5F26

Quant Time: Aug 27 03:13:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 27 03:12:26 2022
Response via : Initial Calibration

