

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW082624\
 Data File : VW037048.D
 Acq On : 26 Aug 2024 23:36
 Operator : SY/MD
 Sample : P3691-03
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K45

Quant Time: Aug 27 04:18:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 27 03:16:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	586534	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	565983	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	264837	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	228646	42.615	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.240%
7) Chloroethane-d5	1.532	69	191710	46.232	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.460%
11) 1,1-Dichloroethene-d2	2.056	65	103789	46.024	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.040%
21) 2-Butanone-d5	3.796	46	228507	109.979	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.980%
24) Chloroform-d	4.243	84	417044	49.664	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.320%
26) 1,2-Dichloroethane-d4	4.937	65	274411	51.892	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.780%
32) Benzene-d6	4.953	84	864412	50.565	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.140%
36) 1,2-Dichloropropane-d6	5.986	67	279233	52.151	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.300%
41) Toluene-d8	7.240	98	769516	49.877	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.760%
43) trans-1,3-Dichloroprop...	7.551	79	108208	43.682	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.360%
47) 2-Hexanone-d5	8.024	63	140663	96.259	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.260%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	333753	51.396	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.800%
66) 1,2-Dichlorobenzene-d4	11.554	152	280927	52.964	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082624\
Data File : VV037048.D
Acq On : 26 Aug 2024 23:36
Operator : SY/MD
Sample : P3691-03
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1K45

Quant Time: Aug 27 04:18:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 27 03:16:59 2024
Response via : Initial Calibration

