

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082624\
 Data File : VV037058.D
 Acq On : 27 Aug 2024 03:13
 Operator : SY/MD
 Sample : P3691-13
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K57

Quant Time: Aug 27 04:20:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 27 03:16:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	547960	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	538534	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	255926	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	208988	41.693	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.380%
7) Chloroethane-d5	1.532	69	185383	47.853	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.700%
11) 1,1-Dichloroethene-d2	2.056	65	99061	47.020	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.040%
21) 2-Butanone-d5	3.793	46	216314	111.439	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.440%
24) Chloroform-d	4.243	84	411738	52.484	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.960%
26) 1,2-Dichloroethane-d4	4.937	65	270292	54.711	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.420%
32) Benzene-d6	4.953	84	824483	50.688	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.380%
36) 1,2-Dichloropropane-d6	5.986	67	270561	53.107	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.220%
41) Toluene-d8	7.239	98	731772	49.848	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.700%
43) trans-1,3-Dichloroprop...	7.551	79	103418	43.877	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.760%
47) 2-Hexanone-d5	8.024	63	143851	103.458	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.460%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	339787	54.992	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	109.980%
66) 1,2-Dichlorobenzene-d4	11.554	152	270979	52.867	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082624\
Data File : VV037058.D
Acq On : 27 Aug 2024 03:13
Operator : SY/MD
Sample : P3691-13
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1K57

Quant Time: Aug 27 04:20:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 27 03:16:59 2024
Response via : Initial Calibration

