

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW082624\
 Data File : VW037035.D
 Acq On : 26 Aug 2024 18:07
 Operator : SY/MD
 Sample : P3692-01
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K89

Quant Time: Aug 27 02:03:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 27 01:58:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	651496	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	623931	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	298105	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	245237	41.150	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.300%
7) Chloroethane-d5	1.529	69	202917	44.055	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.120%
11) 1,1-Dichloroethene-d2	2.057	65	110218	44.002	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.000%
21) 2-Butanone-d5	3.796	46	266639	115.535	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	115.540%
24) Chloroform-d	4.243	84	433480	46.474	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.940%
26) 1,2-Dichloroethane-d4	4.937	65	287143	48.885	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.760%
32) Benzene-d6	4.953	84	895888	47.539	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.080%
36) 1,2-Dichloropropane-d6	5.986	67	287644	48.732	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.460%
41) Toluene-d8	7.240	98	802596	47.190	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.380%
43) trans-1,3-Dichloroprop...	7.551	79	128615	47.098	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.200%
47) 2-Hexanone-d5	8.024	63	180470	112.030	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.030%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	363968	50.843	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.680%
66) 1,2-Dichlorobenzene-d4	11.555	152	294550	49.335	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082624\
Data File : VV037035.D
Acq On : 26 Aug 2024 18:07
Operator : SY/MD
Sample : P3692-01
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1K89

Quant Time: Aug 27 02:03:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 27 01:58:15 2024
Response via : Initial Calibration

