

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082624\
 Data File : VV037042.D
 Acq On : 26 Aug 2024 20:41
 Operator : SY/MD
 Sample : P3693-25
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K92

Quant Time: Aug 27 02:06:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 27 01:58:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	612287	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	587590	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	271710	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	232940	41.589	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.180%
7) Chloroethane-d5	1.532	69	196571	45.410	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.820%
11) 1,1-Dichloroethene-d2	2.056	65	103708	44.054	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.100%
21) 2-Butanone-d5	3.796	46	249720	115.133	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	115.130%
24) Chloroform-d	4.243	84	417310	47.606	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.220%
26) 1,2-Dichloroethane-d4	4.937	65	277853	50.333	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.660%
32) Benzene-d6	4.953	84	866625	48.831	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.660%
36) 1,2-Dichloropropane-d6	5.986	67	281409	50.625	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.240%
41) Toluene-d8	7.239	98	775686	48.428	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.860%
43) trans-1,3-Dichloroprop...	7.551	79	117570	45.716	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.440%
47) 2-Hexanone-d5	8.024	63	159997	105.463	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.460%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	339768	50.398	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.800%
66) 1,2-Dichlorobenzene-d4	11.554	152	283826	52.157	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082624\
Data File : VV037042.D
Acq On : 26 Aug 2024 20:41
Operator : SY/MD
Sample : P3693-25
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1K92

Quant Time: Aug 27 02:06:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 27 01:58:15 2024
Response via : Initial Calibration

