

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082720\
 Data File : VV018107.D
 Acq On : 27 Aug 2020 11:11
 Operator : SY/MD
 Sample : VV0827MBL01
 Misc : 5.0uL/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK96

Quant Time: Aug 28 06:28:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 28 06:20:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	277269	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	272394	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	132851	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	95370	45.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.16%
7) Chloroethane-d5	1.58	69	85668	49.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.00%
11) 1,1-Dichloroethene-d2	2.12	63	146660	36.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.48%
21) 2-Butanone-d5	3.91	46	124604	89.93	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.93%
24) Chloroform-d	4.37	84	179597	46.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.28%
26) 1,2-Dichloroethane-d4	5.05	65	120582	48.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.42%
32) Benzene-d6	5.07	84	362108	48.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.48%
36) 1,2-Dichloropropane-d6	6.09	67	117712	47.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.92%
41) Toluene-d8	7.33	98	321914	47.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.36%
43) trans-1,3-Dichloropropene-	7.64	79	54197	48.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.38%
47) 2-Hexanone-d5	8.11	63	63965	81.22	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.22%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	137928	43.16	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.32%
64) 1,2-Dichlorobenzene-d4	11.65	152	135387	51.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.56%

Target Compounds					Ovalue
25) Chloroform	4.40	83	5452	1.390 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082720\
Data File : VV018107.D
Acq On : 27 Aug 2020 11:11
Operator : SY/MD
Sample : VV0827MBL01
Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK96

Quant Time: Aug 28 06:28:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 28 06:20:16 2020
Response via : Initial Calibration

