

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082720\
 Data File : VV018108.D
 Acq On : 27 Aug 2020 11:34
 Operator : SY/MD
 Sample : VV0827WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK97

Quant Time: Aug 28 06:30:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 28 06:20:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	263570	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	262603	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	128299	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	87342	43.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.86%
7) Chloroethane-d5	1.58	69	78648	47.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.60%
11) 1,1-Dichloroethene-d2	2.12	63	135191	35.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.28%
21) 2-Butanone-d5	3.91	46	120820	91.73	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.73%
24) Chloroform-d	4.38	84	168714	46.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.18%
26) 1,2-Dichloroethane-d4	5.06	65	117158	49.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.54%
32) Benzene-d6	5.08	84	343844	47.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.02%
36) 1,2-Dichloropropane-d6	6.09	67	112131	47.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.78%
41) Toluene-d8	7.34	98	301452	46.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.64%
43) trans-1,3-Dichloropropene-	7.64	79	52932	49.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.64%
47) 2-Hexanone-d5	8.11	63	59548	78.43	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.43%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	135176	43.87	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.74%
64) 1,2-Dichlorobenzene-d4	11.65	152	130373	51.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.26%

Target Compounds					Ovalue
25) Chloroform	4.40	83	5161	1.385 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082720\
Data File : VV018108.D
Acq On : 27 Aug 2020 11:34
Operator : SY/MD
Sample : VV0827WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK97

Quant Time: Aug 28 06:30:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 28 06:20:16 2020
Response via : Initial Calibration

