

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082721\
 Data File : VV021969.D
 Acq On : 27 Aug 2021 12:27
 Operator : SY/MD
 Sample : VV0827MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK241

Quant Time: Aug 28 02:39:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 28 02:34:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	381786	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	365684	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	185549	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	130829	49.483	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.960%
7) Chloroethane-d5	1.564	69	101895	49.351	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.700%
11) 1,1-Dichloroethene-d2	2.105	63	164045	37.517	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.040%
21) 2-Butanone-d5	3.876	46	192582	106.209	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.210%
24) Chloroform-d	4.349	84	225389	46.752	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.500%
26) 1,2-Dichloroethane-d4	5.034	65	143179	50.422	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.840%
32) Benzene-d6	5.053	84	498046	50.213	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.420%
36) 1,2-Dichloropropane-d6	6.072	67	156509	50.175	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.360%
41) Toluene-d8	7.317	98	467202	49.899	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.800%
43) trans-1,3-Dichloroprop...	7.625	79	71151	47.925	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.840%
47) 2-Hexanone-d5	8.088	63	118986	99.416	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.420%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	191770	48.004	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.000%
66) 1,2-Dichlorobenzene-d4	11.628	152	191453	52.486	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.980%
Target Compounds						
					Qvalue	
16) Methylene chloride	2.510	84	3089	1.047	ug/L	98
25) Chloroform	4.384	83	12058	2.402	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082721\
Data File : VV021969.D
Acq On : 27 Aug 2021 12:27
Operator : SY/MD
Sample : VV0827MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK241

Quant Time: Aug 28 02:39:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 28 02:34:14 2021
Response via : Initial Calibration

