

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082721\
 Data File : VV021970.D
 Acq On : 27 Aug 2021 12:51
 Operator : SY/MD
 Sample : VV0827WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK242

Quant Time: Aug 28 02:39:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 28 02:34:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	371490	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	354961	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	184403	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	126866	49.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.62%
7) Chloroethane-d5	1.564	69	100375	49.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.92%
11) 1,1-Dichloroethene-d2	2.105	63	159373	37.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.92%
21) 2-Butanone-d5	3.876	46	181045	102.61	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.61%
24) Chloroform-d	4.352	84	213523	45.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.04%
26) 1,2-Dichloroethane-d4	5.034	65	141641	51.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.52%
32) Benzene-d6	5.053	84	486020	50.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.96%
36) 1,2-Dichloropropane-d6	6.072	67	150921	49.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.70%
41) Toluene-d8	7.320	98	448986	49.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.80%
43) trans-1,3-Dichloroprop...	7.625	79	70376	48.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.66%
47) 2-Hexanone-d5	8.088	63	110596	95.20	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.20%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	186386	48.07	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.14%
66) 1,2-Dichlorobenzene-d4	11.628	152	186809	51.53	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.06%
Target Compounds						
25) Chloroform	4.375	83	16534	3.38	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082721\
Data File : VV021970.D
Acq On : 27 Aug 2021 12:51
Operator : SY/MD
Sample : VV0827WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK242

Quant Time: Aug 28 02:39:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 28 02:34:14 2021
Response via : Initial Calibration

