

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082721\
 Data File : VV021970.D
 Acq On : 27 Aug 2021 12:51
 Operator : SY/MD
 Sample : VV0827WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Aug 28 02:39:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 28 02:34:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	371490	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	354961	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	184403	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	126866	49.31	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	98.62%		
7) Chloroethane-d5	1.564	69	100375	49.96	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	99.92%		
11) 1,1-Dichloroethene-d2	2.105	63	159373	37.46	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	74.92%		
21) 2-Butanone-d5	3.876	46	181045	102.61	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	102.61%		
24) Chloroform-d	4.352	84	213523	45.52	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.04%		
26) 1,2-Dichloroethane-d4	5.034	65	141641	51.26	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	102.52%		
32) Benzene-d6	5.053	84	486020	50.48	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	100.96%		
36) 1,2-Dichloropropane-d6	6.072	67	150921	49.85	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	99.70%		
41) Toluene-d8	7.320	98	448986	49.40	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	98.80%		
43) trans-1,3-Dichloroprop...	7.625	79	70376	48.83	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	97.66%		
47) 2-Hexanone-d5	8.088	63	110596	95.20	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	95.20%		
56) 1,1,2,2-Tetrachloroeth...	10.220	84	186386	48.07	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	96.14%		
66) 1,2-Dichlorobenzene-d4	11.628	152	186809	51.53	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	103.06%		
Target Compounds						
25) Chloroform	4.375	83	16534	3.38	ug/L	Qvalue 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082721\
Data File : VV021970.D
Acq On : 27 Aug 2021 12:51
Operator : SY/MD
Sample : VV0827WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Aug 28 02:39:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 28 02:34:14 2021
Response via : Initial Calibration





#25
Chloroform
Concen: 3.38 ug/L
RT: 4.375 min Scan# 1037
Delta R.T. -0.003 min
Lab File: VV021970.D
Acq: 27 Aug 2021 12:51

Instrument :
MSVOA_V
ClientSampleId :

Tgt Ion: 83 Resp: 16534
Ion Ratio Lower Upper
83 100
85 70.2 46.3 85.9

