

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW082823\
 Data File : VW031946.D
 Acq On : 28 Aug 2023 16:24
 Operator : SY/MD
 Sample : 04175-05ME
 Misc : 6.08g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EZYX0ME

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/31/2023
 Supervised By :Mahesh Dadoda 08/31/2023

Quant Time: Aug 29 00:08:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 26 00:50:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	161385	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	155524	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	79762	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	94937	93.997	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 188.000%#			
7) Chloroethane-d5	1.526	69	57464	64.903	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 129.800%			
11) 1,1-Dichloroethene-d2	2.040	65	46112	97.502	ug/L	-0.02
Spiked Amount 50.000	Range 60 - 125		Recovery = 195.000%#			
21) 2-Butanone-d5	3.809	46	228772	287.993	ug/L	0.01
Spiked Amount 100.000	Range 40 - 130		Recovery = 287.990%#			
24) Chloroform-d	4.249	84	263825	119.507	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 239.020%#			
26) 1,2-Dichloroethane-d4	4.941	65	178319	121.716	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 243.440%#			
32) Benzene-d6	4.957	84	515448	121.459	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 242.920%#			
36) 1,2-Dichloropropane-d6	5.989	67	183988	134.697	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 269.400%#			
41) Toluene-d8	7.246	98	458578	117.276	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 234.560%#			
43) trans-1,3-Dichloroprop...	7.555	79	91095	128.670	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 257.340%#			
47) 2-Hexanone-d5	8.066	63	115243m	275.270	ug/L	0.04
Spiked Amount 100.000	Range 45 - 130		Recovery = 275.270%#			
56) 1,1,2,2-Tetrachloroeth...	10.162	84	259640	147.096	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 294.200%#			
66) 1,2-Dichlorobenzene-d4	11.567	152	187998	129.295	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 258.600%#			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082823\
Data File : VV031946.D
Acq On : 28 Aug 2023 16:24
Operator : SY/MD
Sample : 04175-05ME
Misc : 6.08g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EZYX0ME

Quant Time: Aug 29 00:08:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 26 00:50:15 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 08/31/2023
Supervised By : Mahesh Dadoda 08/31/2023

