

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV082918\  
 Data File : VV007253.D  
 Acq On : 29 Aug 2018 13:06  
 Operator : SY/MD  
 Sample : J4649-11  
 Misc : 25.0 mL/MSVOA V/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 GAB94

Quant Time: Aug 30 04:55:48 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR082818WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Thu Aug 30 04:42:02 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.67  | 114  | 208568   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 8.90  | 117  | 212198   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.30 | 152  | 82962    | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.32   | 65    | 72497    | 4.90     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 98.00%  |
| 7) Chloroethane-d5             | 1.58   | 69    | 65783    | 5.51     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 110.20% |
| 11) 1,1-Dichloroethene-d2      | 2.13   | 63    | 99564    | 3.89     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 77.80%  |
| 20) 2-Butanone-d5              | 3.97   | 46    | 221636   | 59.96    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 119.92% |
| 24) Chloroform-d               | 4.41   | 84    | 163911   | 5.48     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 109.60% |
| 26) 1,2-Dichloroethane-d4      | 5.09   | 65    | 83379    | 5.59     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 111.80% |
| 32) Benzene-d6                 | 5.10   | 84    | 304239   | 5.04     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 100.80% |
| 36) 1,2-Dichloropropane-d6     | 6.12   | 67    | 104614   | 5.33     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 106.60% |
| 41) Toluene-d8                 | 7.37   | 98    | 225989   | 4.31     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 86.20%  |
| 43) trans-1,3-Dichloropropene- | 7.67   | 79    | 31801    | 4.56     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 91.20%  |
| 46) 2-Hexanone-d5              | 8.15   | 63    | 69789    | 38.07    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 76.14%  |
| 57) 1,1,2,2-Tetrachloroethane- | 10.27  | 84    | 79279    | 5.20     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 104.00% |
| 64) 1,2-Dichlorobenzene-d4     | 11.68  | 152   | 88387    | 5.55     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 111.00% |

| Target Compounds |      |    |       |           | Ovalue |
|------------------|------|----|-------|-----------|--------|
| 13) Acetone      | 2.21 | 43 | 10315 | 5.56 ug/L | 88     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV082918\  
Data File : VV007253.D  
Acq On : 29 Aug 2018 13:06  
Operator : SY/MD  
Sample : J4649-11  
Misc : 25.0 mL/MSVOA V/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
GAB94

Quant Time: Aug 30 04:55:48 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR082818WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Thu Aug 30 04:42:02 2018  
Response via : Initial Calibration

